

User Instructions for Commodes

T304 – T306 – T317 – T58 – T59 – T60

Introduction

Congratulations on your purchase of your Cefndy Healthcare Commode. Please ensure you read and understand these instructions in full before construction and use. Please retain these for future reference.

The stackable commode is designed for user comfort. There is a heat sealed padded upholstered back which can be cleaned without risk of water ingress.

The commodes all have non-marking, non-slip feet, come with a 5L lidded potty and have a 25mm modesty cover.

There are several commode options with different features. See feature list below.

	T304	T306	T317	T58	T59	T60
Seat height	450 – 600mm	455mm	405 – 550mm	470 – 620mm	455mm	450 – 600mm
Width Between Arms	470mm	470mm	470mm	495mm	470mm	470mm
Toilet seat Depth	420mm	420mm	420mm	355mm	420mm	420mm
Seat to top of Arms	180mm	180mm	180mm	191mm	180mm	180mm
Seat to top of Back	330mm	330mm	330mm	330mm	330mm	330mm
Footprint (max)	465 x 500mm	465 x 500mm	455 x 490mm	460 x 515mm	465 x 500mm	465 x 500mm
Maximum User Weight	159kg 25st	159kg 25st	159kg 25st	159kg 25st	159kg 25st	159kg 25st



T58 – Adj. Height & detachable arms.



T306 – Fixed Height, blue frame & upholstery.



T60 - Adjustable Height

T304 - Adjustable Height, blue frame & upholstery.

T317 – Adjustable Reduced Height

Pack Contents

- Commode
- User Instructions

Assembly Instructions

Please note:

- When unpacking carefully examine the product to ensure it is complete.
- Before use check for any damage. If damaged please contact your supplier and refrain from using the product.
- **Do not attempt to use if the product is damaged.**

Correct Use

- Remove the vinyl modesty cover to use the commode.
- Replace the vinyl modesty cover note: - this is not designed to be sat on as it is not fixed to the frame it is for cosmetic purposes only.
- Please empty the potty at regular intervals and be aware of its maximum capacity.
- Do not attempt to move the commode whilst occupied

Locating the Potty:

- To locate the potty in the commode, remove the loose modesty cover and raise the toilet seat to a vertical position resting it against the backrest.
- By holding the potty handle in the vertical position, place the potty, lid uppermost on the support brackets ensuring an accurate location of the wide brim on the brackets.
- Remove lid, fold the potty handle down and lower the toilet seat into the horizontal position.

Removing the Potty:

- Once the commode has been used, repeat the process in the reverse order ensuring that the lid is correctly placed to avoid possible spillage.

Height Adjustment – T58, T60, T304 & T317 Only.

- Remove the pin clip from each of the legs.
- Select the required height of the frame by gently sliding each extension leg in or out.
- When desired height is found refit the pin clip into the appropriate hole in each extension leg.
- Ensuring it goes fully through. Ensure that all legs are set to the same height.

Detachable Arms (T58 Only)

- Using both hands remove each commode arm by lifting squarely from their fixing sockets. To refit the arms, carefully locate both ends into the fixing sockets and push fully down.

Cleaning

- This product can be cleaned using a mild detergent and a mild disinfectant or antibacterial solution. The framework should be wiped down on at least a weekly basis. Caution must be taken to understand the suitability of cleaning agents/chemicals being used.
- Do not use an abrasive cleaner as this may damage the frame and invalidate your warranty.
- It is recommended that the potty is thoroughly cleaned after every use, using an appropriate cleaning agent.

Use of power/jet washing & industrial washing machines

- If either of the above is used to clean equipment care must be taken since high-pressure water can be penetrative.
- Care should be taken that frames are not placed too close to high-pressure water jets.
- After cleaning ensure any residual water is removed from the framework as it could cause permanent rusting of the frame (leave frames open in a warm environment for several hours to ensure all internal moisture evaporates).
- We cannot offer advice on industrial washing machines with regards to specific machine settings, as these will be dependent on the manufacturer's instructions.

Maintenance Requirements

- The product should be checked on a regular basis for damage or wear and tear – this includes flaking or cracking paint. Any undue signs of damage should be noted and reported to the supplier in the first instance. Should damage be evident, then the user should refrain from using the product until advice has been sought from the supplier.

Warning

- This product is intended for light use i.e. to aid dining, reading or writing etc.
– It is not designed to support heavy loads.
- We recommend a full risk assessment is carried out before use. This should include consideration for the equipment's intended use, the capacity of the user to safely use the equipment, the environment in which the equipment is intended to be used and **all** other relevant considerations.

Warranty

This product carries a 12 month warranty from the date of purchase against manufacturing defects. Misuse or damage is not covered. This warranty does not affect your statutory rights. As manufacturers we cannot specify expected product life as this is based on the level of usage.

Disposal Instructions

Appropriate hygiene standards should be applied before disposal to minimise the risk of infection.

- Materials used are essentially non-hazardous and can be disposed of appropriately consideration should be given to local re-cycling arrangements.

Contact Details

We have attempted to cover most eventualities, however if you have any specific queries, please contact:

Cefndy Healthcare
Cefndy Road
Rhyl
Denbighshire
LL18 2HG

Tel: 01824 706005

E-mail: cefndy.sales@denbighshire.gov.uk