

User Instructions

March 2008

Introduction to the M2 Gas-Tip Shower, Toilet & Commode Chair



The M2 Gas-Tip Shower, Toilet and Commode Chair provides definitive positioning and postural management for those who are unable to maintain their body posture and require extra support when showering or toileting.

The gas spring-operated

tilt-in-space enables optimal positioning and redistribution of pressure whilst showering and toileting. Tilting the body has the effect of distributing weight over a larger contact surface area providing a simple and effective way to combat the build up of pressure and change pressure points.

As with most of the shower, toilet and commode chairs they are manufactured from high quality stainless steel that will not degrade over time. In addition, the M2 Gas-Tip has moulded polyurethane upholstery to ensure exceptional comfort, easy cleaning and disinfection.

Why is there a need for this type of chair?

The ability of a seated person to function efficiently and perform activities depends on their ability to adopt the appropriate posture. All activity is posture dependent. If a person cannot move or has problems adjusting their posture, it may be necessary to use seating to try to provide externally what is limited internally.

The principles of good seating are equally applicable to showering, toileting and commode chairs as they are to general seating. Whether looking for basic shower stools, practical shower, toileting and commode chairs that help to maintain independence or dynamic, postural support shower, toileting and commode chairs for assisted bathing, the Kirton $\rm H_20$ collection can accommodate both user and carer needs.

Who should consider using the M2 Gas-Tip Shower, Toilet & Commode Chair?

For those with or at risk of developing an unconventional body shape, those who have postural-related problems or those who are unable to maintain their body posture. For domestic home use, hospital, hospice and care home environments.

Maximum weight limit 140kg/22stones.

M2 Gas-Tip Shower, Toilet & Commode Chair

Features

- Manual, gas spring-operated tilt-in-space
 - Adjustable seat height
- Open, horseshoe seat with seat cover
- Cushioned polyurethane armrests
- Lockable, swivel armrests for side transference
- Height adjustable footrests
- Supportive calf rest
- Flexible, adjustable net-weave backrest
- Height adjustable headrest
- Lateral supports
- 125mm wheels, braked
- Fits over standard toilet
- Attendant propelled
- Stainless steel framework, powder lacquered

Optional accessories

- Commode Pan
- · Height and depth adjustable headrest
- Standard aperture seat
- Legrest elevation kit
- Footrest plate for mounting on legrest elevation kit
- Drip stand mounting kit
- Hipstrap

Questions and Concerns

Should you experience any difficulties or have any concerns regarding the initial purchase or subsequently adjusting the M2 Gas-Tip Shower, Toilet & Commode chair or have any concerns regarding its use or operation then immediately contact either your Kirton representative, distributor or our Customer Service Team on Freephone 0800 212709 or +44 (0) 1440 705352, they will be delighted to help you.

Guidelines for Best Practice

To ensure good practice when considering purchasing an M2 Gas-Tip Shower, Toilet & Commode chair and effective post purchase we advise that the following guidelines are considered:

- Seek advice and input from a qualified therapist/professional and/or Kirton representative/distributor at the initial prescription of the chair.
- If unsure when ordering the chair and any accessories contact either the therapist involved or your Kirton representative, distributor or our Customer Service Team on Freephone 0800 212709 or +44 (0) 1440 705352, to discuss further.
- Users of the M2 Gas-Tip Shower, Toilet & Commode chair, where necessary, should be supervised to ensure the most appropriate position is maintained in the chair.



User Instructions

March 2008

The User Instructions

The purpose of the following user instructions is to provide a guide on how to utilise, adjust and maintain the M2 Gas-Tip Shower, Toilet & Commode chair.

The instructions should be read by all those involved with the supervision of the user and/or the care of the chair before use.

Please retain these instructions with the Care and Maintenance card enclosed for future reference.

Wheels



Before using the M2 Gas-Tip chair for showering or toileting ensure the brakes are applied on all four wheels by pressing the grey ribbed area of the wheels down.

Apply all brakes when tilting the chair.

Tilting the chair



The gas action tilts the chair either forwards or backwards. Before moving the chair ensure the user's arms are positioned on the armrests or placed in the user's lap. To tilt, hold the push handles, squeeze the lever and tilt the chair forwards or backwards.

When hoisting a user into the chair a reclined tilt position is recommended to enable positioning of the user into the back of the seat area from the hoist sling.



Care should be taken when using the negative tilt angle as the user could slide out of the chair. The forward tilt can be restricted by using the block attached at the base of the frame enabling the chair to only move to a horizontal position.

Adjusting the seat height



The M2 Gas-Tip chair has telescopic legs enabling the seat height to be adjusted to suit the user.

Apply the brakes to all four wheels.

Lay the chair on its side unoccupied and remove the height adjustment clips by pushing on the two open sides of the clips with your thumbs.

Slide the leg in and out to the desired height and replace the clips. Make sure the clips are pushed all the way through the legs to ensure stability.

Repeat the procedure on the remaining three legs.

All wheels should be at the same height.

Footrests



The footrests slot into the two holders at the front of the chair frame.

The height of the footrests can be adjusted to suit individual users. Remove the height adjustment clips by

pushing on the two open sides of the clips with your thumbs. Slide the footrest in and out to the desired height and replace the clips ensuring they are all the way through the stainless steel tubing.

The angle of the footplates can be inclined. Simply remove each footrest from the chair, detach the adjustment clips, turn the footplate 180° and replace the clips. Position the footrests on the opposite side of the chair.

The calf support must always be used with the M2 Gas-Tip to prevent entrapment of the user's legs behind the footrests when tilted.



The calf support is fastened around the footrests with Velcro. The tension of the calf support can be adjusted according to the user's needs.



User Instructions

March 2008

Armrests



The armrests have the ability to be swiveled backwards to facilitate side transfers by

lifting the snap release. NB: if using the lateral supports these will need to be removed before lifting the armrest.

The armrests can be removed completely by lifting the snap release, vertically lifting the armrest and pulling free of the chair over the positioning bolt. The armrests should always be secured in a horizontal position when tilting the chair for the safety and positioning of the user.

Backrest Adjustment



The tension of the backrest can be adjusted to accommodate the user's positioning and postural requirements by pulling the webbing straps tighter or releasing them accordingly.



The backrest can be completely removed for cleaning by pressing the two plastic tabs together and pulling the buckle apart. Washable at 80 °c maximum.

Commode Pan



If using a commode pan, ensure the pan is slid straight forwards and backwards along the seat. On the underside of the seat module the commode pan holders can be tightened by

bending the wings closer to the seat to enable a firmer grip on the commode pan or indeed loosened accordingly.

Lateral Supports



The lateral supports are positioned in the square support holders on each side of the chair.



The width of each lateral support is adjusted by loosening the thumbscrew, positioning the support and then retightening.



The height of the lateral supports can be changed by swapping over the left and the right supports and positioning them in their alternative holders.

Headrest



The M2 Gas-Tip chair has a height adjustable headrest. To adjust the height, loosen the thumbscrew on the holder at the base of the headrest, reposition and retighten.



The wings of the headrest can be moved in and out to support the head. Apply pressure to either side of the headrest and push inwards to reduce the interior size of the headrest or alternatively position your hands on the inside of the headrest and


pull out to increase the interior size of the headrest. DO NOT APPLY PRESSURE TO THE TWO ENDS OF THE HEADREST AS IT WILL BREAK THE EDGES.



User Instructions

March 2008

Removing and replacing the seat module



The seat can be removed and replaced for cleaning or if multiple seats are purchased for individuals for use with one frame.

The seat module is removed by lifting the seat module upwards from the frame holding both sides of the seat module. The seat is replaced by applying a light pressure on either side of the seat module. Ensure the back edge of the seat is fitted around the flat metal bar at the back of the chair.

The seat lid fits over the aperture and covers the entire seat module.

Potential Risks, Cautions and Warnings

- The user is at risk if the M2 Gas-Tip chair is not properly set up to their requirements.
- Ensure the seat module is correctly "snapped" back in place before use.
- Before using the chair for showering or toileting apply all four brakes.
- Do not stand upright on the footrests.
- Always use the calf support, particularly when tilting the chair backwards.
- Do not apply pressure to the two ends of the headrest.
- Ensure the thumbscrews on the headrest and lateral supports are tight before use.
- Do not sit on the arms of the chair.
- Ensure the armrests are locked in a horizontal position before tilting the chair.
- Ensure the backrest tension is adjusted to avoid undue pressure on the back.
- Be aware that when using the negative tilt facility the user may slide forward and potentially out of the chair.

Care and Maintenance

- Wipe down the framework with soapy water and a soft cloth to reduce the build up of soap and limescale.
- Cleaning with a scouring pad is not advised.
- Maximum temperature 80 ℃.

Helpline

Should you have any concerns either pre or post purchase of the chair contact our Customer Service Team on Freephone 0800 212709 or +44 (0) 1440 705352, to discuss further. Alternatively, if you purchased your chair from a Kirton distributor in the first instance please contact your point of reference there.



23 Rookwood Way Haverhill Suffolk, CB9 8PB England Telephone: +44 (0) 1440 705352

Fax: +44 (0) 1440 706199
Freephone: 0800 212709
Email: info@kirtonhealthcare.co.uk
www.kirton-healthcare.co.uk





